



EPI-Lo Instrument Calibration Plan

Matt Hill, Don Mitchell

Matthew.Hill@jhuapl.edu

Don.Mitchell@jhuapl.edu



EPI-Lo Calibration

- EPI-Lo measurements are intended to generate the information needed to derive differential intensities ($j[\text{cm}^2 \text{-sr-s-keV}]^{-1}$)
- The goal of EPI-Lo characterization and calibration efforts, is to develop the quantitative procedures for converting the count rates (R [counts s^{-1}]) reported by EPI-Lo into estimates of j for the various defined ranges of energies, particle species, and arrival angles
- “Calibration” for a particle instrument like EPI-Lo means determining the following:
 - Transfer function from counts into flux (physical units)
 - Characteristic of “Rate-out” versus “Rate-in”
 - Response to visible and ultraviolet light
 - Response to high energetic particle backgrounds



Calibration Types

Foreground

Ions and Electrons in the energy range of interest to the instrument

Background

Electrons

Characterize the rates from penetrating radiation

Light

Characterize rejection of UV background, primarily H-alpha

Characterize rejection of sunlight and glint



Accuracy & Precision

- **Accuracy and precision requirement (Level IV):**
 - **30% accuracy**
 - **20% Cross-Instrument precision (reduced to 10% in flight)**

Parameter	Required	Goal (Capability)	Comment/Heritage
Electron Energies	50 - 500keV	25 - 1000 keV	Electron capability from JEDI, RBSPICE
Ion Energies	50 keV/nucleon – 15000 keV Total E	50 keV/nucleon – 15000 keV Total E	Capability partially based on RBSPICE capabilities. Top energy ~250keV/nuc for Fe
Energy Resolution	45% for required energy range	40% for required energy range	Telemetry limited
Time sampling	5 sec	1 sec	Telemetry and/or statistics limited
Angle resolution	<30° x <30°	Ions, ~15° x 12° to <30° x <30° e-, 45°	Varies with elevation
Pitch Angle (PA) Coverage	0°-90° or 90°-180°, some samples in both hemispheres	0°-90° or 90°-180°, some samples in both hemispheres	
Time for Full PA	1 – 5 sec	1 – 5 sec	Telemetry limited
Ion Composition	H, He3, He4, C, O, Ne, Mg, Si, Fe	H, He3, He4, C, O, Ne, Mg, Si, Fe	He3/He4 ~50 to 1000 keV/nuc
Electron Sensitivity: I=Intensity (1/cm ² .sr)	j = 1E1-1E6/cm ² -s-sr	Sensor-G:0.144 (cm ² .sr) Pixel-G: ~0.02 (cm ² .sr) Up to 6E6 1/s counting	j=Intensity (1/cm ² -s-sr) G=Geom. Factor (cm ² -sr) 8 pixels/sensor
Ion Sensitivity	j = 1E1-1E6/cm ² -s-sr	Sensor-G:0.16 (cm ² .sr) Pixel-G: ~0.002 (cm ² .sr) Up to 3.5E6 1/s counting (TOFxE)	80 pixels/sensor



Foreground Calibration Requirements

Foreground Electrons

– **40 keV to 1000 keV (Needed for understanding backgrounds)**

- 1 keV to 30 keV Electron Gun at APL
- 30 keV to 100 keV Radioactive sources at APL
- 125 keV to 1.6 MeV Accelerator at GSFC

Foreground Ions (H, He3, He4, O, Fe)

– **40 keV to 15000 keV (Level 4 Requirements)**

– **(Goal: protons to 15 MeV)**

- 3 keV to 170 keV Accelerator at APL
- 30 keV to 5 MeV Degraded alpha sources
- 125 keV to 1.6 MeV Accelerator at GSFC
- 1 MeV to 15 MeV Accelerator at LBL



Test as You Fly

- The instruments will be tested in flight-like environments
- Since the instrument will need to operate in a high background environment, we will characterize response to high energy penetrating radiation, UV light, and low energy plasma (all potential sources of background counts for EPI-Lo)



Background Calibration Requirement

Background Electrons

– 1 keV to 10 MeV (from the expected environment)

- 1 keV to 50 keV Electron Gun at APL
- 125 keV to 1.6 MeV Accelerator at GSFC
- 1 MeV to 10 MeV Accelerator at Idaho

Background Ions

– 3 keV to 50 MeV (from the expected environment)

- 3 keV to 170 keV Accelerator at APL
- 30 keV to 5 MeV Degraded alpha sources
- 125 keV to 1.6 MeV Accelerator at GSFC
- 1 MeV to ~100 MeV Accelerator at LBL

Photons

- UV and visible lamps at APL




The JHU/APL Calibration Facility


- The APL particle accelerator is a versatile system capable of producing a broad range of ion species at energies from 20 to 170 keV
 - The system includes a electron-impact ionization source, extraction gap, Einzel Lens and Wien filter mounted in the insulated terminal structure along with all associated power supplies
 - The system will produce beams of H, He, O, and noble gas ions with intensities over the range of 100' s to 1,000,000 particles/cm²/sec at the target position (mm² - cm²)
 - We also have a variety of radioactive sources as stimulus.
-



The JHU/APL Calibration Facility



- All ions from a gas source
- Energy continuously tunable: 3 to 170 kV
- Wien filter
- Beam intensity between 10 and 10^{10} ions/cm²/s
- Purposed built articulation stage






Beam Tests at GSFC

It is planned to have four calibration sessions at the accelerator at the Goddard Space Flight Center. Each session starts with a one-day setup, check, and pump:

- Session 1: Exploratory run to characterize EPI-Lo
- Session 2: Use H beam to scan both angles to complete characterization of the transfer function
- Session 3: Characterize sensor response with e- beam from ~ 100 keV to 1 MeV
- Session 4: Use heavy ions (He, O, and Ar) to characterize the instrument response




GSFC JEDI EM calibration results





GSFC JEDI EM calibration results



Essentially the same figure with log scales and a different histogramming code.

Ghost peaks have appropriate TOF for beam, but low energy. They only appear for high energy heavies, and so are not expected to significantly affect foreground measurements of lower energy particles (they will be a minor contribution to “noise”)



EPI-Lo Development Summary

- Component Testing
 - Aliveness
 - Limited Performance Test
- Assembly Testing
 - Functional
 - Limited Performance Test
- Prototype “EM” Testing
 - Validate instrument design and performance
 - Energy response;
 - Instrument efficiency;
 - Instrument geometry factor
 - Establish testing procedures
- Flight Model (FM)
 - Qualified each unit as proto-flight unit
 - Test to acceptance levels



Calibration Approach (Flight units)

- FM Unit
 - All instrument integration activities will be performed in a Class 5 clean room environment
 - Test in bell jar to characterize geometry, energy response, and sensitivity
 - Compare with EM results
 - Calibrate in Beam Facilities at APL, GSFC, and LBL to characterize energy response, sensitivity, dynamic range
- Test Philosophy:
 - FM will be fully calibrated
 - pre- and post-environmental qualification spot calibration
 - In-flight cross calibration between EPI-Lo and EPI-Hi



EPI-Lo Calibration Schedule

- Final calibration for FM slated for three weeks
- Major calibration efforts will be performed at APL facility
 - Substantial understandings of the instrument performance will be made with the EPI-Lo EM characteristics



Tests at Instrument Level

	Element	Property	Requirement*	Expected Performance	Calibration Accuracy
System Level	Calibration mode without Collimator	Energy -ToF plane characteristic	Verify simulation to 20%	Verify simulation to 5%	5%
		Input/output rate at system level	Known to 10%	Known to 2%	1%
		Background rejection	> 90%	> 95%	2%
		Mass resolution	Discriminate between e^- , H^+ , He^+ , CNO^+	< 1.5 AMU (H^+) < 4 AMU (CNO) < 12 AMU (Fe^+)	1 AMU
		Absolute efficiency	Known to 50% for e^- , H^+ , He^+ , NO^+	10%	10%
	Calibration mode with collimator	Scattering of ions	< 0.5%	< 0.1%	0.1%
		Scattering of electrons	< 0.5%	< 0.01%	0.1%
		Properties at sector boundaries	Known to 30%	Known to 5%	5%
		Efficiency as a function of azimuth	Known to 20%	Known to 5%	5%
		Angular resolution	30°	25°	3°
		Geometric factor	> 0.03 cm ² -sr	0.03 cm ² -sr	0.01 cm ² -sr
		Full calibration: verify previous measurements			
	Flight mode	Input/output rate at system level	Known to 50% 30	Known to 10%	10%
		Verify all modes			
		Verify all timing windows			
		Throughput of event classification			
		Efficiency of counters			
		Energy -ToF plane characteristics			
		Threshold settings			
		Temperature dependent			

*Science requires relative/absolute accuracy: 20%/50%. Ground calibration 20% precision, reduced to 10% in flight.



In-Flight Calibration

- On-orbit and cruise calibration achieves relative calibration to 10% precision
 - Uniformity confirmed by evolution of pitch angle distribution from onset to shock passage.
 - Such calibrations cover the entire energy and FoV coverage with linear instrument response (targeted rates, no pulse pileup)
- Built-in features to determine on-orbit instrument ion performance
 - *Measure pulse-height spectrum of secondary electrons from incident protons as function of time-of-flight*
 - *Unit has built-in stimulus to inject known pulse through the front-end electronics*



Summary

- Calibration plan satisfies all Level IV requirements
- Calibration facilities have been identified that meet EPI-Lo needs
- APL operates and maintains the key EPI-Lo calibration facility which allows maximum flexibility
- Begin component testing
- Questions?